

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
 Data File : VY011162.D
 Acq On : 28 Oct 2022 17:18
 Operator : KP/MD
 Sample : N5316-07
 Misc : 12.89g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOTAL-VOC-GRAB-7-NE-QUARTER

Quant Time: Oct 28 23:46:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	221236	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	379393	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	323215	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	117327	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	157354	68.785	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 137.580%	
35) Dibromofluoromethane	7.715	113	149229	63.961	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 127.920%	
50) Toluene-d8	10.178	98	543445	58.795	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 117.580%	
62) 4-Bromofluorobenzene	12.477	95	148525	49.109	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 98.220%	
Target Compounds						Qvalue
20) Methylene Chloride	4.710	84	28089	6.248	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
Data File : VY011162.D
Acq On : 28 Oct 2022 17:18
Operator : KP/MD
Sample : N5316-07
Misc : 12.89g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOTAL-VOC-GRAB-7-NE-QUARTER

Quant Time: Oct 28 23:46:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration

