

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
 Data File : VY011163.D
 Acq On : 28 Oct 2022 17:41
 Operator : KP/MD
 Sample : N5316-13
 Misc : 8.85g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOTAL-VOC-GRAB-1-PILE-2

Quant Time: Oct 28 23:46:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	286308	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	498875	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	435317	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	184533	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	189368	63.965	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.940%
35) Dibromofluoromethane	7.716	113	187151	61.003	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.000%
50) Toluene-d8	10.179	98	717161	59.006	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	118.020%
62) 4-Bromofluorobenzene	12.477	95	213562	53.701	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
Data File : VY011163.D
Acq On : 28 Oct 2022 17:41
Operator : KP/MD
Sample : N5316-13
Misc : 8.85g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOTAL-VOC-GRAB-1-PILE-2

Quant Time: Oct 28 23:46:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration

