

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
 Data File : VY011165.D
 Acq On : 28 Oct 2022 18:26
 Operator : KP/MD
 Sample : N5332-01
 Misc : 6.57g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-4-WC

Quant Time: Oct 28 23:47:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	195601	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	354543	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	303292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	116252	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	150327	74.326	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 148.660%	
35) Dibromofluoromethane	7.716	113	140123	64.268	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 128.540%	
50) Toluene-d8	10.179	98	517255	59.884	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 119.760%	
62) 4-Bromofluorobenzene	12.477	95	142381	50.377	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.760%	
Target Compounds						
20) Methylene Chloride	4.716	84	49871	20.421	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
Data File : VY011165.D
Acq On : 28 Oct 2022 18:26
Operator : KP/MD
Sample : N5332-01
Misc : 6.57g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-4-WC

Quant Time: Oct 28 23:47:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration

