

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
 Data File : VY011172.D
 Acq On : 28 Oct 2022 22:13
 Operator : KP/MD
 Sample : VY1028SBL02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1028SBL02

Quant Time: Oct 28 23:48:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	250765	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	440626	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	393440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	175037	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	135967	52.437	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.880%
35) Dibromofluoromethane	7.716	113	132403	48.863	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.720%
50) Toluene-d8	10.179	98	507443	47.270	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.540%
62) 4-Bromofluorobenzene	12.477	95	156633	44.593	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
Data File : VY011172.D
Acq On : 28 Oct 2022 22:13
Operator : KP/MD
Sample : VY1028SBL02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1028SBL02

Quant Time: Oct 28 23:48:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration

