

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
 Data File : VY011175.D
 Acq On : 28 Oct 2022 23:43
 Operator : KP/MD
 Sample : N5261-02
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC

Quant Time: Oct 29 00:47:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	247859	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	432725	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	367060	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	150622	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	153722	59.980	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 119.960%	
35) Dibromofluoromethane	7.716	113	153444	57.662	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 115.320%	
50) Toluene-d8	10.179	98	546557	51.844	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.680%	
62) 4-Bromofluorobenzene	12.477	95	160084	46.407	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 92.820%	
Target Compounds						
20) Methylene Chloride	4.716	84	36107	8.320	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
Data File : VY011175.D
Acq On : 28 Oct 2022 23:43
Operator : KP/MD
Sample : N5261-02
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Time: Oct 29 00:47:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration

