

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
 Data File : VY011178.D
 Acq On : 29 Oct 2022 00:51
 Operator : KP/MD
 Sample : N5327-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LACY-05

Quant Time: Oct 29 01:41:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	29269	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	45585	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	37905	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	12420	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	19391	64.071	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.140%	
35) Dibromofluoromethane	7.716	113	19295	68.830	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 137.660%	
50) Toluene-d8	10.179	98	60326	54.319	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.640%	
62) 4-Bromofluorobenzene	12.477	95	14649	40.312	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 80.620%	
Target Compounds						
					Qvalue	
16) Acetone	3.942	43	1150	15.530	ug/l #	81
20) Methylene Chloride	4.710	84	6742	17.695	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
Data File : VY011178.D
Acq On : 29 Oct 2022 00:51
Operator : KP/MD
Sample : N5327-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LACY-05

Quant Time: Oct 29 01:41:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration

