

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
 Data File : VY011179.D
 Acq On : 29 Oct 2022 01:14
 Operator : KP/MD
 Sample : N5327-04
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LACY-03

Quant Time: Oct 29 01:41:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	162527	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	298375	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	263583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	102231	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	129151	76.850	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	153.700%
35) Dibromofluoromethane	7.716	113	120387	65.610	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	131.220%
50) Toluene-d8	10.179	98	394175	54.225	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.440%
62) 4-Bromofluorobenzene	12.477	95	113518	47.726	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.460%
Target Compounds						
20) Methylene Chloride	4.710	84	45206	22.987	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
Data File : VY011179.D
Acq On : 29 Oct 2022 01:14
Operator : KP/MD
Sample : N5327-04
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LACY-03

Quant Time: Oct 29 01:41:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration

