

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
 Data File : VY011184.D
 Acq On : 29 Oct 2022 03:07
 Operator : KP/MD
 Sample : N5310-08
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BIN 1-2-(2840)-VOC-4

Quant Time: Oct 29 06:47:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	23329	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	38820	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	35266	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	13094	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	18406	76.302	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	152.600%
35) Dibromofluoromethane	7.716	113	15924	66.703	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	133.400%
50) Toluene-d8	10.179	98	51244	54.183	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.360%
62) 4-Bromofluorobenzene	12.477	95	14533	46.962	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.920%
Target Compounds						
20) Methylene Chloride	4.716	84	3621	9.376	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
Data File : VY011184.D
Acq On : 29 Oct 2022 03:07
Operator : KP/MD
Sample : N5310-08
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BIN 1-2-(2840)-VOC-4

Quant Time: Oct 29 06:47:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration

