

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
 Data File : VY011186.D
 Acq On : 29 Oct 2022 03:53
 Operator : KP/MD
 Sample : N5310-10
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BIN 1-2-(2840)-VOC-6

Quant Time: Oct 29 06:47:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	9941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	17295	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	15634	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	5120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	8612	83.781	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 167.560%#	
35) Dibromofluoromethane	7.716	113	5209	48.976	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 97.960%	
50) Toluene-d8	10.179	98	23162	54.970	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.940%	
62) 4-Bromofluorobenzene	12.477	95	6200	44.970	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 89.940%	
Target Compounds						
20) Methylene Chloride	4.704	84	1767	11.871	ug/l	Qvalue # 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
Data File : VY011186.D
Acq On : 29 Oct 2022 03:53
Operator : KP/MD
Sample : N5310-10
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BIN 1-2-(2840)-VOC-6

Quant Time: Oct 29 06:47:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration

