

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
 Data File : VY011180.D
 Acq On : 29 Oct 2022 01:37
 Operator : KP/MD
 Sample : N5286-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BBS-SOILCUT-1026

Quant Time: Oct 29 06:46:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.789	168	142167	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	243048	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	217116	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	88427	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	94553	64.320	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.640%	
35) Dibromofluoromethane	7.716	113	93420	62.503	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 125.000%	
50) Toluene-d8	10.179	98	328998	55.561	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 111.120%	
62) 4-Bromofluorobenzene	12.477	95	93893	48.461	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 96.920%	
Target Compounds						
20) Methylene Chloride	4.710	84	41125	24.220	ug/l	Qvalue # 82


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
Data File : VY011180.D
Acq On : 29 Oct 2022 01:37
Operator : KP/MD
Sample : N5286-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BBS-SOILCUT-1026

Quant Time: Oct 29 06:46:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY011180.D
Acq: 29 Oct 2022 01:37




Instrument :
MSVOA_Y
ClientSampleId :
BBS-SOILCUT-1026


Tgt Ion:168 Resp: 142167
Ion Ratio Lower Upper
168 100
99 55.0 46.9 70.3



#20
Methylene Chloride
Concen: 24.220 ug/l
RT: 4.710 min Scan# 474
Delta R.T. -0.006 min
Lab File: VY011180.D
Acq: 29 Oct 2022 01:37

Tgt Ion: 84 Resp: 41125
Ion Ratio Lower Upper
84 100
49 113.0 114.4 171.6#
51 35.5 35.3 52.9
86 61.6 53.7 80.5





#33

1,2-Dichloroethane-d4

Concen: 64.320 ug/l

RT: 8.143 min Scan# 1103

Delta R.T. 0.006 min

Lab File: VY011180.D

Acq: 29 Oct 2022 01:37

Instrument :

MSVOA_Y

ClientSampleId :





BBS-SOILCUT-1026

Tgt Ion: 65 Resp: 94553

Ion Ratio Lower Upper

65 100

67 54.9 0.0 99.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY011180.D

Acq: 29 Oct 2022 01:37




Tgt Ion: 114 Resp: 243048


Ion Ratio Lower Upper

114 100

63 18.8 0.0 43.2

88 14.9 0.0 31.8









#35
Dibromofluoromethane
Concen: 62.503 ug/l
RT: 7.716 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY011180.D
Acq: 29 Oct 2022 01:37




Instrument :
MSVOA_Y
ClientSampleId :
BBS-SOILCUT-1026


Tgt Ion:113 Resp: 93420
Ion Ratio Lower Upper
113 100
111 102.6 82.6 124.0
192 19.3 17.6 26.4



#50
Toluene-d8
Concen: 55.561 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY011180.D
Acq: 29 Oct 2022 01:37

Tgt Ion: 98 Resp: 328998
Ion Ratio Lower Upper
98 100
100 64.6 50.2 75.2









#62
4-Bromofluorobenzene
Concen: 48.461 ug/l
RT: 12.477 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY011180.D
Acq: 29 Oct 2022 01:37




Instrument :
MSVOA_Y
ClientSampleId :
BBS-SOILCUT-1026


Tgt Ion: 95 Resp: 93893
Ion Ratio Lower Upper
95 100
174 84.0 0.0 177.0
176 79.8 0.0 175.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 1586
Delta R.T. -0.000 min
Lab File: VY011180.D
Acq: 29 Oct 2022 01:37


Tgt Ion: 117 Resp: 217116
Ion Ratio Lower Upper
117 100
82 54.5 43.0 64.4
119 32.1 25.4 38.0





#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.422 min Scan# 1903
 Delta R.T. 0.000 min
 Lab File: VY011180.D
 Acq: 29 Oct 2022 01:37

Instrument :
 MSVOA_Y
 ClientSampleId :
 BBS-SOILCUT-1026



Tgt Ion:152 Resp: 88427
 Ion Ratio Lower Upper
 152 100
 115 57.2 28.2 84.7
 150 156.6 0.0 347.6

