

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
 Data File : VY011189.D
 Acq On : 29 Oct 2022 05:01
 Operator : KP/MD
 Sample : N5310-13
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BIN 1-2-(2840)-VOC-9

Quant Time: Oct 29 06:48:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.795	168	454	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	875	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	406	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	80	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	547	116.521	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	233.040%#
35) Dibromofluoromethane	7.710	113	164	30.478	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	60.960%
50) Toluene-d8	10.185	98	941	44.142	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.280%
62) 4-Bromofluorobenzene	12.483	95	152	21.791	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	43.580%
Target Compounds						Qvalue
20) Methylene Chloride	4.729	84	48	3.897	ug/l #	39
76) 1,2,3-Trichloropropane	12.483	75	20	24.046	ug/l #	100
81) trans-1,4-Dichloro-2-b...	12.483	75	20	51.066	ug/l #	7


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
Data File : VY011189.D
Acq On : 29 Oct 2022 05:01
Operator : KP/MD
Sample : N5310-13
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BIN 1-2-(2840)-VOC-9

Quant Time: Oct 29 06:48:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY011189.D
Acq: 29 Oct 2022 05:01




Instrument :
MSVOA_Y
ClientSampleId :
BIN 1-2-(2840)-VOC-9


Tgt Ion:168 Resp: 454
Ion Ratio Lower Upper
168 100
99 56.8 46.9 70.3



#20
Methylene Chloride
Concen: 3.897 ug/l
RT: 4.729 min Scan# 477
Delta R.T. 0.013 min
Lab File: VY011189.D
Acq: 29 Oct 2022 05:01

Tgt Ion: 84 Resp: 48
Ion Ratio Lower Upper
84 100
49 82.6 114.4 171.6#
51 0.0 35.3 52.9#
86 0.0 53.7 80.5#





#33

1,2-Dichloroethane-d4

Concen: 116.521 ug/l

RT: 8.136 min Scan# 1036

Delta R.T. -0.001 min

Lab File: VY011189.D

Acq: 29 Oct 2022 05:01

Instrument :

MSVOA_Y

ClientSampleId :





BIN 1-2-(2840)-VOC-9

Tgt Ion: 65 Resp: 547

Ion Ratio Lower Upper

65 100

67 51.4 0.0 99.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY011189.D

Acq: 29 Oct 2022 05:01




Tgt Ion: 114 Resp: 875


Ion Ratio Lower Upper

114 100

63 15.9 0.0 43.2

88 13.3 0.0 31.8









#35
Dibromofluoromethane
Concen: 30.478 ug/l
RT: 7.710 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY011189.D
Acq: 29 Oct 2022 05:01




Instrument :
MSVOA_Y
ClientSampleId :
BIN 1-2-(2840)-VOC-9


Tgt Ion:113 Resp: 164
Ion Ratio Lower Upper
113 100
111 272.6 82.6 124.0#
192 14.6 17.6 26.4#



#50
Toluene-d8
Concen: 44.142 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. 0.006 min
Lab File: VY011189.D
Acq: 29 Oct 2022 05:01

Tgt Ion: 98 Resp: 941
Ion Ratio Lower Upper
98 100
100 57.5 50.2 75.2









#62
4-Bromofluorobenzene
Concen: 21.791 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.006 min
Lab File: VY011189.D
Acq: 29 Oct 2022 05:01




Instrument :
MSVOA_Y
ClientSampleId :
BIN 1-2-(2840)-VOC-9


Tgt Ion:	95	Resp:	152
Ion	Ratio	Lower	Upper
95	100		
174	94.1	0.0	177.0
176	85.5	0.0	175.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.483 min Scan# 1585
Delta R.T. -0.007 min
Lab File: VY011189.D
Acq: 29 Oct 2022 05:01

Tgt Ion:	117	Resp:	406
Ion	Ratio	Lower	Upper
117	100		
82	44.1	43.0	64.4
119	0.0	25.4	38.0









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.434 min Scan# 1749
Delta R.T. 0.012 min
Lab File: VY011189.D
Acq: 29 Oct 2022 05:01




Instrument :
MSVOA_Y
ClientSampleId :
BIN 1-2-(2840)-VOC-9


Tgt Ion:152 Resp: 80
Ion Ratio Lower Upper
152 100
115 56.3 28.2 84.7
150 292.5 0.0 347.6



#76
1,2,3-Trichloropropane
Concen: 24.046 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.147 min
Lab File: VY011189.D
Acq: 29 Oct 2022 05:01


Tgt Ion: 75 Resp: 20
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#81
 trans-1,4-Dichloro-2-butene
 Concen: 51.066 ug/l
 RT: 12.483 min Scan# 11
 Delta R.T. 0.109 min
 Lab File: VY011189.D
 Acq: 29 Oct 2022 05:01

Instrument :
 MSVOA_Y
 ClientSampleId :
 BIN 1-2-(2840)-VOC-9



Tgt Ion: 75 Resp: 20
 Ion Ratio Lower Upper
 75 100
 53 0.0 85.3 127.9#
 89 0.0 37.0 55.4#

