

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
 Data File : VY020037.D
 Acq On : 28 Oct 2024 13:52
 Operator : SY/MD
 Sample : P4575-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PL-02-102424

Quant Time: Oct 29 01:57:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	51719	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	100003	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	85701	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	25540	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	36574	57.433	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.860%
35) Dibromofluoromethane	7.640	113	33842	51.283	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.560%
50) Toluene-d8	10.109	98	118373	48.574	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.140%
62) 4-Bromofluorobenzene	12.408	95	31864	36.187	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	72.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
Data File : VY020037.D
Acq On : 28 Oct 2024 13:52
Operator : SY/MD
Sample : P4575-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PL-02-102424

Quant Time: Oct 29 01:57:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

