

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
 Data File : VY020038.D
 Acq On : 28 Oct 2024 14:16
 Operator : SY/MD
 Sample : P4574-04
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 GRAVEL-2

Quant Time: Oct 29 01:58:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	195248	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	398098	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	346273	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	117047	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	134499	55.946	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.900%
35) Dibromofluoromethane	7.634	113	131773	50.161	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.320%
50) Toluene-d8	10.109	98	489414	50.449	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.900%
62) 4-Bromofluorobenzene	12.402	95	139672	39.846	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	79.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
Data File : VY020038.D
Acq On : 28 Oct 2024 14:16
Operator : SY/MD
Sample : P4574-04
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
GRAVEL-2

Quant Time: Oct 29 01:58:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

