

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
 Data File : VY020039.D
 Acq On : 28 Oct 2024 15:15
 Operator : SY/MD
 Sample : P4574-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 GRAVEL-1

Quant Time: Oct 29 01:58:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	137514	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	286119	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	251393	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	85077	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	93145	55.011	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.020%
35) Dibromofluoromethane	7.634	113	94378	49.986	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.980%
50) Toluene-d8	10.109	98	350289	50.239	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.480%
62) 4-Bromofluorobenzene	12.408	95	103909	41.245	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	82.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
Data File : VY020039.D
Acq On : 28 Oct 2024 15:15
Operator : SY/MD
Sample : P4574-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
GRAVEL-1

Quant Time: Oct 29 01:58:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

