

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
 Data File : VY020042.D
 Acq On : 28 Oct 2024 16:28
 Operator : SY/MD
 Sample : P4561-07
 Misc : 5.16g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BP-F-18-VOC

Quant Time: Oct 29 01:59:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	59608	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	116923	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	103231	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	33342	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	44366	60.448	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 120.900%	
35) Dibromofluoromethane	7.634	113	40851	52.946	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.900%	
50) Toluene-d8	10.109	98	140385	49.270	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 98.540%	
62) 4-Bromofluorobenzene	12.408	95	40431	39.271	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 78.540%	
Target Compounds						
16) Acetone	3.866	43	4690	24.547	ug/l	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
Data File : VY020042.D
Acq On : 28 Oct 2024 16:28
Operator : SY/MD
Sample : P4561-07
Misc : 5.16g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BP-F-18-VOC

Quant Time: Oct 29 01:59:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

