

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
 Data File : VY020043.D
 Acq On : 28 Oct 2024 17:05
 Operator : SY/MD
 Sample : P4567-11
 Misc : 5.87g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-3-VOC

Quant Time: Oct 29 02:00:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	180459	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	378580	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	351313	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	121818	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	127086	57.195	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.380%
35) Dibromofluoromethane	7.634	113	125289	50.151	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.300%
50) Toluene-d8	10.109	98	476467	51.646	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.300%
62) 4-Bromofluorobenzene	12.407	95	143832	43.148	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	86.300%
Target Compounds						
16) Acetone	3.879	43	22199	38.379	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
Data File : VY020043.D
Acq On : 28 Oct 2024 17:05
Operator : SY/MD
Sample : P4567-11
Misc : 5.87g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-3-VOC

Quant Time: Oct 29 02:00:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

