

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
 Data File : VY020029.D
 Acq On : 28 Oct 2024 09:29
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 10/29/2024
 Supervised By :Mahesh Dadoda 10/29/2024

Quant Time: Oct 29 01:51:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	206029	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	8.627	114	357320	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.420	117	306579	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	147527	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	126771	49.972	ug/l	0.01
Spiked Amount 50.000	Range 50 - 163		Recovery =	99.940%		
35) Dibromofluoromethane	7.646	113	127235	53.961	ug/l	0.01
Spiked Amount 50.000	Range 54 - 147		Recovery =	107.920%		
50) Toluene-d8	10.115	98	455263	52.284	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	104.560%		
62) 4-Bromofluorobenzene	12.413	95	163597	51.997	ug/l	0.00
Spiked Amount 50.000	Range 29 - 146		Recovery =	104.000%		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.873	85	79697	41.553	ug/l	100
3) Chloromethane	2.080	50	122651	49.134	ug/l	99
4) Vinyl Chloride	2.214	62	137508	50.058	ug/l	99
5) Bromomethane	2.610	94	92611	53.264	ug/l	96
6) Chloroethane	2.744	64	99637	53.992	ug/l	98
7) Trichlorofluoromethane	3.074	101	205170	50.195	ug/l	98
8) Diethyl Ether	3.464	74	57797	46.525	ug/l	71
9) 1,1,2-Trichlorotrifluo...	3.830	101	121950	51.293	ug/l	92
10) Methyl Iodide	4.019	142	112317	46.709	ug/l	93
11) Tert butyl alcohol	4.878	59	37778	178.815	ug/l	# 74
12) 1,1-Dichloroethene	3.805	96	107007	48.413	ug/l	# 81
13) Acrolein	3.671	56	19104	102.486	ug/l	99
14) Allyl chloride	4.403	41	211745	53.095	ug/l	# 90
15) Acrylonitrile	5.073	53	143646	252.807	ug/l	97
16) Acetone	3.884	43	189367	286.756	ug/l	# 81
17) Carbon Disulfide	4.122	76	245136	41.377	ug/l	99
18) Methyl Acetate	4.403	43	77619	54.563	ug/l	# 86
19) Methyl tert-butyl Ether	5.128	73	318882	50.218	ug/l	95
20) Methylene Chloride	4.640	84	127839	52.431	ug/l	# 86
21) trans-1,2-Dichloroethene	5.128	96	121169	50.369	ug/l	87
22) Diisopropyl ether	6.030	45	496829	57.195	ug/l	# 90
23) Vinyl Acetate	5.976	43	1284239	257.885	ug/l	# 91
24) 1,1-Dichloroethane	5.927	63	272014	56.333	ug/l	96
25) 2-Butanone	6.908	43	230515	260.783	ug/l	# 83
26) 2,2-Dichloropropane	6.902	77	234914	54.627	ug/l	97
27) cis-1,2-Dichloroethene	6.902	96	159171	53.542	ug/l	81
28) Bromochloromethane	7.256	49	110668	52.071	ug/l	# 72
29) Tetrahydrofuran	7.274	42	121587	240.906	ug/l	# 81
30) Chloroform	7.433	83	281519	57.193	ug/l	99
31) Cyclohexane	7.713	56	205663	48.520	ug/l	86
32) 1,1,1-Trichloroethane	7.628	97	242695	56.261	ug/l	94
36) 1,1-Dichloropropene	7.847	75	184456	54.300	ug/l	94
37) Ethyl Acetate	6.994	43	85793	48.434	ug/l	# 95
38) Carbon Tetrachloride	7.829	117	202338	55.562	ug/l	96
39) Methylcyclohexane	9.115	83	216299	50.535	ug/l	# 84
40) Benzene	8.091	78	570603	55.502	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
 Data File : VY020029.D
 Acq On : 28 Oct 2024 09:29
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 10/29/2024
 Supervised By :Mahesh Dadoda 10/29/2024

Quant Time: Oct 29 01:51:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.231	41	53145	55.435	ug/l #	82
42) 1,2-Dichloroethane	8.170	62	159618	54.043	ug/l	94
43) Isopropyl Acetate	8.207	43	181886	50.350	ug/l #	88
44) Trichloroethene	8.877	130	133402	53.575	ug/l	94
45) 1,2-Dichloropropane	9.152	63	146668	56.998	ug/l	96
46) Dibromomethane	9.237	93	74041	52.665	ug/l	89
47) Bromodichloromethane	9.432	83	214099	57.373	ug/l	94
48) Methyl methacrylate	9.231	41	82213	51.738	ug/l #	80
49) 1,4-Dioxane	9.237	88	14980	952.008	ug/l #	84
51) 4-Methyl-2-Pentanone	10.005	43	475210	257.904	ug/l	88
52) Toluene	10.176	92	362556	56.497	ug/l	99
53) t-1,3-Dichloropropene	10.402	75	182389	54.321	ug/l	97
54) cis-1,3-Dichloropropene	9.865	75	217550	54.911	ug/l #	86
55) 1,1,2-Trichloroethane	10.578	97	99521	53.729	ug/l	93
56) Ethyl methacrylate	10.444	69	137463	51.720	ug/l #	75
57) 1,3-Dichloropropane	10.725	76	176829	54.361	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.719	63	342145	276.525	ug/l	92
59) 2-Hexanone	10.767	43	348423	264.753	ug/l	85
60) Dibromochloromethane	10.920	129	127848	53.776	ug/l	98
61) 1,2-Dibromoethane	11.023	107	85281	50.950	ug/l	99
64) Tetrachloroethene	10.652	164	108297	49.663	ug/l	92
65) Chlorobenzene	11.450	112	385750	55.011	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.523	131	135040	56.920	ug/l	98
67) Ethyl Benzene	11.523	91	722890	56.751	ug/l	99
68) m/p-Xylenes	11.633	106	519871	110.508	ug/l	89
69) o-Xylene	11.962	106	248010	54.869	ug/l	90
70) Styrene	11.974	104	433757	56.879	ug/l	96
71) Bromoform	12.139	173	66575	52.419	ug/l #	95
73) Isopropylbenzene	12.261	105	703143	56.664	ug/l	98
74) N-amyl acetate	12.078	43	160282	48.768	ug/l #	86
75) 1,1,2,2-Tetrachloroethane	12.511	83	115061	52.184	ug/l	99
76) 1,2,3-Trichloropropane	12.560	75	66959m	23.169	ug/l	
77) Bromobenzene	12.541	156	136772	51.940	ug/l	84
78) n-propylbenzene	12.602	91	861161	57.434	ug/l	96
79) 2-Chlorotoluene	12.682	91	482023	56.075	ug/l	93
80) 1,3,5-Trimethylbenzene	12.743	105	556655	55.547	ug/l	95
81) trans-1,4-Dichloro-2-b...	12.310	75	36908	51.548	ug/l #	84
82) 4-Chlorotoluene	12.785	91	487706	55.488	ug/l	95
83) tert-Butylbenzene	13.005	119	516248	57.504	ug/l	94
84) 1,2,4-Trimethylbenzene	13.047	105	552717	55.630	ug/l	95
85) sec-Butylbenzene	13.182	105	768178	57.270	ug/l	97
86) p-Isopropyltoluene	13.297	119	613226	56.301	ug/l	96
87) 1,3-Dichlorobenzene	13.291	146	284252	53.454	ug/l	97
88) 1,4-Dichlorobenzene	13.371	146	278219	53.256	ug/l	97
89) n-Butylbenzene	13.620	91	606658	57.042	ug/l	98
90) Hexachloroethane	13.883	117	125555	59.200	ug/l	83
91) 1,2-Dichlorobenzene	13.663	146	244629	52.355	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.279	75	16531	46.978	ug/l	75
93) 1,2,4-Trichlorobenzene	14.925	180	127439	48.641	ug/l	98
94) Hexachlorobutadiene	15.029	225	70632	48.490	ug/l	97
95) Naphthalene	15.151	128	228866	46.315	ug/l	99
96) 1,2,3-Trichlorobenzene	15.334	180	105414	47.596	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
Data File : VY020029.D
Acq On : 28 Oct 2024 09:29
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024

Quant Time: Oct 29 01:51:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.719 min Scan# 99
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 206029
Ion Ratio Lower Upper
168 100
99 64.6 39.1 58.7




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#2
Dichlorodifluoromethane
Concen: 41.553 ug/l
RT: 1.873 min Scan# 25
Delta R.T. 0.007 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion: 85 Resp: 79697
Ion Ratio Lower Upper
85 100
87 32.1 16.0 48.0









#3
Chloromethane
Concen: 49.134 ug/l
RT: 2.080 min Scan# 50
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 50 Resp: 12265
Ion Ratio Lower Upper
50 100
52 31.3 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#4
Vinyl Chloride
Concen: 50.058 ug/l
RT: 2.214 min Scan# 81
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29


Tgt Ion: 62 Resp: 137508
Ion Ratio Lower Upper
62 100
64 31.9 26.1 39.1





#5
Bromomethane
Concen: 53.264 ug/l
RT: 2.610 min Scan# 144
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 94 Resp: 9261
Ion Ratio Lower Upper
94 100
96 88.3 74.1 111.1

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024




#6
Chloroethane
Concen: 53.992 ug/l
RT: 2.744 min Scan# 168
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29



Tgt Ion: 64 Resp: 99637
Ion Ratio Lower Upper
64 100
66 31.3 26.1 39.1









#7
Trichlorofluoromethane
Concen: 50.195 ug/l
RT: 3.074 min Scan# 21
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 205170
Ion Ratio Lower Upper
101 100
103 63.4 49.3 73.9




Manual Integrations
APPROVED

Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#8
Diethyl Ether
Concen: 46.525 ug/l
RT: 3.464 min Scan# 286
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion: 74 Resp: 57797
Ion Ratio Lower Upper
74 100
45 108.9 41.3 124.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.293 ug/l
RT: 3.830 min Scan# 344
Delta R.T. 0.018 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 121950
Ion Ratio Lower Upper
101 100
85 49.2 34.5 51.7
151 70.4 61.7 92.5




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#10
Methyl Iodide
Concen: 46.709 ug/l
RT: 4.019 min Scan# 377
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion:142 Resp: 112317
Ion Ratio Lower Upper
142 100
127 46.9 32.8 49.2
141 14.6 11.4 17.2





#11

Tert butyl alcohol

Concen: 178.815 ug/l

RT: 4.878 min Scan# 514

Delta R.T. 0.018 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 59 Resp: 37778

Ion Ratio Lower Upper

59 100





57 12.2 3.0 4.6

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024

Supervised By :Mahesh Dadoda 10/29/2024



#12

1,1-Dichloroethene

Concen: 48.413 ug/l

RT: 3.805 min Scan# 342

Delta R.T. 0.018 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29




Tgt Ion: 96 Resp: 107007


Ion Ratio Lower Upper

96 100

61 173.4 114.6 171.8#

98 59.3 50.8 76.2





#13

Acrolein

Concen: 102.486 ug/l

RT: 3.671 min Scan# 317

Delta R.T. 0.018 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion: 56 Resp: 19104
 Ion Ratio Lower Upper
 56 100
 55 69.0 55.7 83.5

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024
 Supervised By :Mahesh Dadoda 10/29/2024



#14

Allyl chloride

Concen: 53.095 ug/l




RT: 4.403 min Scan# 440


Delta R.T. 0.018 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29


Tgt Ion: 41 Resp: 211745
 Ion Ratio Lower Upper
 41 100
 39 76.5 58.3 87.5
 76 35.4 39.0 58.6





#15
Acrylonitrile
Concen: 252.807 ug/l
RT: 5.073 min Scan# 51
Delta R.T. 0.018 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 53 Resp: 143640
Ion Ratio Lower Upper
53 100
52 83.2 64.7 97.1
51 38.3 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024




#16
Acetone
Concen: 286.756 ug/l
RT: 3.884 min Scan# 355
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Tgt Ion: 43 Resp: 189367
Ion Ratio Lower Upper
43 100
58 27.9 31.7 47.5#





#17
Carbon Disulfide
Concen: 41.377 ug/l
RT: 4.122 min Scan# 394
Delta R.T. 0.018 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 245130
Ion Ratio Lower Upper
76 100
78 8.8 6.8 10.2




Manual Integrations
APPROVED

Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#18
Methyl Acetate
Concen: 54.563 ug/l
RT: 4.403 min Scan# 440
Delta R.T. 0.018 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion: 43 Resp: 77619
Ion Ratio Lower Upper
43 100
74 21.7 23.3 34.9#









#19
Methyl tert-butyl Ether
Concen: 50.218 ug/l
RT: 5.128 min Scan# 511
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 73 Resp: 31888
Ion Ratio Lower Upper
73 100
57 23.9 17.3 25.9




Manual Integrations
APPROVED

Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#20
Methylene Chloride
Concen: 52.431 ug/l
RT: 4.640 min Scan# 479
Delta R.T. 0.024 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29


Tgt Ion: 84 Resp: 127839
Ion Ratio Lower Upper
84 100
49 131.0 84.4 126.6#
51 35.0 26.0 39.0
86 62.0 49.5 74.3





#21
trans-1,2-Dichloroethene
Concen: 50.369 ug/l
RT: 5.128 min Scan# 511
Delta R.T. 0.018 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 121169
Ion Ratio Lower Upper
96 100
61 141.4 97.5 146.3
98 60.3 50.5 75.7




Manual Integrations
APPROVED

Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#22
Diisopropyl ether
Concen: 57.195 ug/l
RT: 6.030 min Scan# 707
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29


Tgt Ion: 45 Resp: 496829
Ion Ratio Lower Upper
45 100
43 59.6 42.2 63.2
87 23.9 25.0 37.6#
59 11.1 10.5 15.7





#23
 Vinyl Acetate
 Concen: 257.885 ug/l
 RT: 5.976 min Scan# 696
 Delta R.T. 0.018 min
 Lab File: VY020029.D
 Acq: 28 Oct 2024 09:29




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion: 43 Resp: 1284239
 Ion Ratio Lower Upper
 43 100
 86 9.6 10.6 16.0

Manual Integrations
 APPROVED



Reviewed By :Romaben Patel 10/29/2024
 Supervised By :Mahesh Dadoda 10/29/2024




#24
 1,1-Dichloroethane
 Concen: 56.333 ug/l
 RT: 5.927 min Scan# 690
 Delta R.T. 0.012 min
 Lab File: VY020029.D
 Acq: 28 Oct 2024 09:29




Tgt Ion: 63 Resp: 272014
 Ion Ratio Lower Upper
 63 100
 98 6.4 4.0 12.2
 100 4.1 2.6 8.0





#25
2-Butanone
Concen: 260.783 ug/l
RT: 6.908 min Scan# 851
Delta R.T. 0.018 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 23051
Ion Ratio Lower Upper
43 100
72 24.6 27.3 40.9




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#26
2,2-Dichloropropane
Concen: 54.627 ug/l
RT: 6.902 min Scan# 850
Delta R.T. 0.018 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29


Tgt Ion: 77 Resp: 234914
Ion Ratio Lower Upper
77 100
97 22.6 12.0 36.1





#27
 cis-1,2-Dichloroethene
 Concen: 53.542 ug/l
 RT: 6.902 min Scan# 848
 Delta R.T. 0.012 min
 Lab File: VY020029.D
 Acq: 28 Oct 2024 09:29




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050



Tgt Ion: 96 Resp: 15917
 Ion Ratio Lower Upper
 96 100
 61 159.6 0.0 256.2
 98 63.9 0.0 130.0




Manual Integrations
 APPROVED


Reviewed By :Romaben Patel 10/29/2024
 Supervised By :Mahesh Dadoda 10/29/2024



#28
 Bromochloromethane
 Concen: 52.071 ug/l
 RT: 7.256 min Scan# 908
 Delta R.T. 0.012 min
 Lab File: VY020029.D
 Acq: 28 Oct 2024 09:29


Tgt Ion: 49 Resp: 110668
 Ion Ratio Lower Upper
 49 100
 129 1.8 0.0 4.6
 130 63.7 72.7 109.1#





#29
Tetrahydrofuran
Concen: 240.906 ug/l
RT: 7.274 min Scan# 911
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 42 Resp: 12158
Ion Ratio Lower Upper
42 100
72 39.5 43.0 64.4
71 38.9 40.8 61.2

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024




#30
Chloroform
Concen: 57.193 ug/l
RT: 7.433 min Scan# 937
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Tgt Ion: 83 Resp: 281519
Ion Ratio Lower Upper
83 100
85 64.7 52.2 78.4





#31
Cyclohexane
Concen: 48.520 ug/l
RT: 7.713 min Scan# 911
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 20566
Ion Ratio Lower Upper
56 100
69 31.8 28.7 43.1
84 80.2 77.7 116.5




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#32
1,1,1-Trichloroethane
Concen: 56.261 ug/l
RT: 7.628 min Scan# 969
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion: 97 Resp: 242695
Ion Ratio Lower Upper
97 100
99 63.8 50.3 75.5
61 48.7 32.6 48.8





#33

1,2-Dichloroethane-d4

Concen: 49.972 ug/l

RT: 8.073 min Scan# 1040

Delta R.T. 0.012 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 12677

Ion Ratio Lower Upper

65 100





67 53.0 0.0 109.6

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024

Supervised By :Mahesh Dadoda 10/29/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.627 min Scan# 1133

Delta R.T. 0.012 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29




Tgt Ion:114 Resp: 357320


Ion Ratio Lower Upper

114 100

63 20.9 0.0 35.0

88 14.2 0.0 27.2





#35

Dibromofluoromethane

Concen: 53.961 ug/l

RT: 7.646 min Scan# 91

Delta R.T. 0.012 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:113 Resp: 127235

Ion Ratio Lower Upper

113 100

111 102.1 82.2 123.4





192 16.5 15.9 23.9

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024

Supervised By :Mahesh Dadoda 10/29/2024



#36

1,1-Dichloropropene

Concen: 54.300 ug/l

RT: 7.847 min Scan# 1005

Delta R.T. 0.012 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29




Tgt Ion: 75 Resp: 184456


Ion Ratio Lower Upper

75 100

110 32.2 19.0 57.0


77 30.5 25.3 37.9





#37
Ethyl Acetate
Concen: 48.434 ug/l
RT: 6.994 min Scan# 80
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 8579
Ion Ratio Lower Upper
43 100
61 12.4 12.6 19.0
70 9.3 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#38
Carbon Tetrachloride
Concen: 55.562 ug/l
RT: 7.829 min Scan# 1002
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion:117 Resp: 202338
Ion Ratio Lower Upper
117 100
119 98.8 75.8 113.6
121 30.9 24.0 36.0





#39

Methylcyclohexane

Concen: 50.535 ug/l

RT: 9.115 min Scan# 1212

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:	83	Resp:	21629
Ion	Ratio	Lower	Upper
83	100		
55	84.9	51.5	77.3
98	47.8	40.5	60.7

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024

Supervised By :Mahesh Dadoda 10/29/2024



#40

Benzene

Concen: 55.502 ug/l




RT: 8.091 min Scan# 1045

Delta R.T. 0.012 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Tgt Ion:	78	Resp:	570603
Ion	Ratio	Lower	Upper
78	100		
77	25.1	19.5	29.3









#41
Methacrylonitrile
Concen: 55.435 ug/l
RT: 7.231 min Scan# 904
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 41 Resp: 5314
Ion Ratio Lower Upper
41 100
39 63.9 48.0 72.0
67 66.8 75.1 112.7
52 30.3 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#42
1,2-Dichloroethane
Concen: 54.043 ug/l
RT: 8.170 min Scan# 1058
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion: 62 Resp: 159618
Ion Ratio Lower Upper
62 100
98 8.6 0.0 22.0









#43
Isopropyl Acetate
Concen: 50.350 ug/l
RT: 8.207 min Scan# 1064
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 181880
Ion Ratio Lower Upper
43 100
61 27.7 27.9 41.9
87 11.1 13.1 19.7




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#44
Trichloroethene
Concen: 53.575 ug/l
RT: 8.877 min Scan# 1174
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion:130 Resp: 133402
Ion Ratio Lower Upper
130 100
95 100.0 0.0 188.0





#45

1,2-Dichloropropane

Concen: 56.998 ug/l

RT: 9.152 min Scan# 1218

Delta R.T. 0.012 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 146668

Ion Ratio Lower Upper

63 100





65 29.8 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024

Supervised By :Mahesh Dadoda 10/29/2024



#46

Dibromomethane

Concen: 52.665 ug/l

RT: 9.237 min Scan# 1233

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29




Tgt Ion: 93 Resp: 74041


Ion Ratio Lower Upper

93 100

95 82.4 66.2 99.4

174 81.7 80.9 121.3









#47
Bromodichloromethane
Concen: 57.373 ug/l
RT: 9.432 min Scan# 1265
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 214099
Ion Ratio Lower Upper
83 100
85 67.2 50.1 75.1
127 8.1 7.4 11.0




Manual Integrations
APPROVED

Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#48
Methyl methacrylate
Concen: 51.738 ug/l
RT: 9.231 min Scan# 1232
Delta R.T. 0.012 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion: 41 Resp: 82213
Ion Ratio Lower Upper
41 100
69 81.7 84.1 126.1#
39 64.4 43.6 65.4





#49

1,4-Dioxane

Concen: 952.008 ug/l

RT: 9.237 min Scan# 1233

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion: 88 Resp: 14980

Ion	Ratio	Lower	Upper
88	100		
43	35.3	0.0	0.0
58	73.8	49.3	73.9

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#50

Toluene-d8

Concen: 52.284 ug/l

RT: 10.115 min Scan# 1377




Delta R.T. 0.006 min


Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Tgt Ion: 98 Resp: 455263

Ion	Ratio	Lower	Upper
98	100		
100	64.7	52.0	78.0









#51
4-Methyl-2-Pentanone
Concen: 257.904 ug/l
RT: 10.005 min Scan# 1359
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 475210
Ion Ratio Lower Upper
43 100
58 36.5 35.7 53.5




Manual Integrations
APPROVED

Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#52
Toluene
Concen: 56.497 ug/l
RT: 10.176 min Scan# 1387
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29


Tgt Ion: 92 Resp: 362556
Ion Ratio Lower Upper
92 100
91 172.9 139.0 208.6





#53
 t-1,3-Dichloropropene
 Concen: 54.321 ug/l
 RT: 10.402 min Scan# 1424
 Delta R.T. 0.006 min
 Lab File: VY020029.D
 Acq: 28 Oct 2024 09:29




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion: 75 Resp: 182389
 Ion Ratio Lower Upper
 75 100
 77 32.0 24.1 36.1

Manual Integrations
 APPROVED



Reviewed By :Romaben Patel 10/29/2024
 Supervised By :Mahesh Dadoda 10/29/2024




#54
 cis-1,3-Dichloropropene
 Concen: 54.911 ug/l
 RT: 9.865 min Scan# 1336
 Delta R.T. 0.006 min
 Lab File: VY020029.D
 Acq: 28 Oct 2024 09:29




Tgt Ion: 75 Resp: 217550
 Ion Ratio Lower Upper
 75 100
 77 30.6 25.4 38.0
 39 52.0 29.8 44.6#





#55
1,1,2-Trichloroethane
Concen: 53.729 ug/l
RT: 10.578 min Scan# 1452
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 97 Resp: 9952
Ion Ratio Lower Upper
97 100
83 90.1 67.5 101.3
85 57.5 42.9 64.3
99 64.7 46.6 70.0

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024




#56
Ethyl methacrylate
Concen: 51.720 ug/l
RT: 10.444 min Scan# 1431
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29



Tgt Ion: 69 Resp: 137463
Ion Ratio Lower Upper
69 100
41 70.2 42.3 63.5#
39 41.7 22.9 34.3#





#57

1,3-Dichloropropane

Concen: 54.361 ug/l

RT: 10.725 min Scan# 1475

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 176829

Ion Ratio Lower Upper

76 100





78 32.3 26.2 39.4

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024

Supervised By :Mahesh Dadoda 10/29/2024



#58

2-Chloroethyl Vinyl ether

Concen: 276.525 ug/l

RT: 9.719 min Scan# 1312

Delta R.T. 0.006 min

Lab File: VY020029.D




Acq: 28 Oct 2024 09:29


Tgt Ion: 63 Resp: 342145

Ion Ratio Lower Upper

63 100


106 25.5 23.7 35.5





#59
2-Hexanone
Concen: 264.753 ug/l
RT: 10.767 min Scan# 1482
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 34842
Ion Ratio Lower Upper
43 100
58 51.4 31.4 94.0

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024




#60
Dibromochloromethane
Concen: 53.776 ug/l
RT: 10.920 min Scan# 1509
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29



Tgt Ion:129 Resp: 127848
Ion Ratio Lower Upper
129 100
127 79.9 39.0 117.0









#61
1,2-Dibromoethane
Concen: 50.950 ug/l
RT: 11.023 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:107 Resp: 8528
Ion Ratio Lower Upper
107 100
109 94.0 75.9 113.9




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#62
4-Bromofluorobenzene
Concen: 51.997 ug/l
RT: 12.413 min Scan# 1754
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29


Tgt Ion: 95 Resp: 163597
Ion Ratio Lower Upper
95 100
174 71.5 0.0 175.6
176 71.3 0.0 171.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.420 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:117 Resp: 306579
Ion Ratio Lower Upper
117 100
82 60.9 42.4 63.6
119 32.2 25.9 38.9

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024




#64
Tetrachloroethene
Concen: 49.663 ug/l
RT: 10.652 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29



Tgt Ion:164 Resp: 108297
Ion Ratio Lower Upper
164 100
166 129.1 102.6 154.0
129 107.3 72.0 108.0
131 96.4 70.0 105.0





#65

Chlorobenzene

Concen: 55.011 ug/l

RT: 11.450 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:112 Resp: 385750

Ion Ratio Lower Upper

112 100





114 30.6 26.4 39.6

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024

Supervised By :Mahesh Dadoda 10/29/2024



#66

1,1,1,2-Tetrachloroethane

Concen: 56.920 ug/l

RT: 11.523 min Scan# 1608

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29




Tgt Ion:131 Resp: 135040


Ion Ratio Lower Upper

131 100

133 94.9 47.8 143.3

119 65.5 34.1 102.2





#67

Ethyl Benzene

Concen: 56.751 ug/l

RT: 11.523 min Scan# 1608

Delta R.T. 0.000 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 722890

Ion Ratio Lower Upper

91 100





106 30.1 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024

Supervised By :Mahesh Dadoda 10/29/2024



#68

m/p-Xylenes

Concen: 110.508 ug/l

RT: 11.633 min Scan# 1626

Delta R.T. 0.000 min

Lab File: VY020029.D




Acq: 28 Oct 2024 09:29


Tgt Ion:106 Resp: 519871

Ion Ratio Lower Upper

106 100

91 210.7 156.0 234.0









#69
o-Xylene
Concen: 54.869 ug/l
RT: 11.962 min Scan# 1680
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 248010
Ion Ratio Lower Upper
106 100
91 223.3 103.6 310.8




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#70
Styrene
Concen: 56.879 ug/l
RT: 11.974 min Scan# 1682
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion:104 Resp: 433757
Ion Ratio Lower Upper
104 100
78 51.5 37.0 55.4
103 54.3 44.2 66.2









#71
Bromoform
Concen: 52.419 ug/l
RT: 12.139 min Scan# 1707
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:173 Resp: 6657
Ion Ratio Lower Upper
173 100
175 46.7 25.1 75.3
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.352 min Scan# 1908
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion:152 Resp: 147527
Ion Ratio Lower Upper
152 100
115 62.8 28.2 84.7
150 175.7 0.0 345.6





#73

Isopropylbenzene

Concen: 56.664 ug/l

RT: 12.261 min Scan# 1729

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 70314

Ion Ratio Lower Upper

105 100





120 25.5 13.4 40.1

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024

Supervised By :Mahesh Dadoda 10/29/2024



#74

N-ethyl acetate

Concen: 48.768 ug/l

RT: 12.078 min Scan# 1699

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Tgt Ion: 43 Resp: 160282




Ion Ratio Lower Upper


43 100

70 40.1 41.4 62.2#

55 23.0 23.3 34.9#

61 22.0 22.6 33.8#





#75

1,1,2,2-Tetrachloroethane

Concen: 52.184 ug/l

RT: 12.511 min Scan# 1770

Delta R.T. 0.000 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 11506

Ion Ratio Lower Upper

83 100

131 9.3 5.0 14.9





85 64.8 32.0 96.0

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024

Supervised By :Mahesh Dadoda 10/29/2024



#76

1,2,3-Trichloropropane

Concen: 23.169 ug/l m

RT: 12.560 min Scan# 1778

Delta R.T. 0.000 min

Lab File: VY020029.D




Acq: 28 Oct 2024 09:29


Tgt Ion: 75 Resp: 66959

Ion Ratio Lower Upper

75 100


77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 51.940 ug/l
RT: 12.541 min Scan# 1775
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:156 Resp: 136772
Ion Ratio Lower Upper
156 100
77 209.1 87.1 261.3
158 97.1 48.6 145.9




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#78
n-propylbenzene
Concen: 57.434 ug/l
RT: 12.602 min Scan# 1785
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion: 91 Resp: 861161
Ion Ratio Lower Upper
91 100
120 21.1 11.5 34.4









#79
2-Chlorotoluene
Concen: 56.075 ug/l
RT: 12.682 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 48202
Ion Ratio Lower Upper
91 100
126 31.4 17.8 53.5




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#80
1,3,5-Trimethylbenzene
Concen: 55.547 ug/l
RT: 12.743 min Scan# 1808
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion:105 Resp: 556655
Ion Ratio Lower Upper
105 100
120 46.8 25.1 75.2









#81
trans-1,4-Dichloro-2-butene
Concen: 51.548 ug/l
RT: 12.310 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 36908
Ion Ratio Lower Upper
75 100
53 102.7 66.8 100.2
89 47.9 35.8 53.6




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#82
4-Chlorotoluene
Concen: 55.488 ug/l
RT: 12.785 min Scan# 1815
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion: 91 Resp: 487706
Ion Ratio Lower Upper
91 100
126 32.3 17.6 52.9









#83
tert-Butylbenzene
Concen: 57.504 ug/l
RT: 13.005 min Scan# 1851
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:119 Resp: 516248
Ion Ratio Lower Upper
119 100
91 68.2 31.3 93.8
134 25.5 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#84
1,2,4-Trimethylbenzene
Concen: 55.630 ug/l
RT: 13.047 min Scan# 1858
Delta R.T. 0.000 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29

Tgt Ion:105 Resp: 552717
Ion Ratio Lower Upper
105 100
120 43.2 23.2 69.6





#85

sec-Butylbenzene

Concen: 57.270 ug/l

RT: 13.182 min Scan# 1878

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 768178

Ion Ratio Lower Upper

105 100





134 18.9 10.1 30.1

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024

Supervised By :Mahesh Dadoda 10/29/2024



#86

p-Isopropyltoluene

Concen: 56.301 ug/l

RT: 13.297 min Scan# 1899

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29




Tgt Ion:119 Resp: 613226


Ion Ratio Lower Upper

119 100

134 25.4 13.1 39.3

91 25.1 10.7 32.1









#87
 1,3-Dichlorobenzene
 Concen: 53.454 ug/l
 RT: 13.291 min Scan# 1898
 Delta R.T. 0.000 min
 Lab File: VY020029.D
 Acq: 28 Oct 2024 09:29

Instrument : MSVOA_Y
 ClientSampleId : VSTDCCC050

Tgt Ion:146 Resp: 28425

Ion	Ratio	Lower	Upper
146	100		
111	42.4	19.0	57.0
148	63.7	31.7	95.1




Manual Integrations
APPROVED
 Reviewed By :Romaben Patel 10/29/2024
 Supervised By :Mahesh Dadoda 10/29/2024




#88
 1,4-Dichlorobenzene
 Concen: 53.256 ug/l
 RT: 13.371 min Scan# 1911
 Delta R.T. 0.000 min
 Lab File: VY020029.D
 Acq: 28 Oct 2024 09:29

Tgt Ion:146 Resp: 278219

Ion	Ratio	Lower	Upper
146	100		
111	40.6	18.4	55.4
148	63.7	32.4	97.2





#89

n-Butylbenzene

Concen: 57.042 ug/l

RT: 13.620 min Scan# 1951

Delta R.T. 0.000 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion: 91 Resp: 606658

Ion	Ratio	Lower	Upper
91	100		
92	53.1	26.6	79.7
134	23.5	12.9	38.7

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024



#90

Hexachloroethane

Concen: 59.200 ug/l

RT: 13.883 min Scan# 1995




Delta R.T. 0.000 min


Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Tgt Ion:117 Resp: 125555


Ion	Ratio	Lower	Upper
117	100		
201	60.3	37.4	112.2





#91
1,2-Dichlorobenzene
Concen: 52.355 ug/l
RT: 13.663 min Scan# 1958
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 244629
Ion Ratio Lower Upper
146 100
111 43.4 19.9 59.7
148 64.0 31.8 95.4

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 10/29/2024
Supervised By :Mahesh Dadoda 10/29/2024




#92
1,2-Dibromo-3-Chloropropane
Concen: 46.978 ug/l
RT: 14.279 min Scan# 2060
Delta R.T. 0.006 min
Lab File: VY020029.D
Acq: 28 Oct 2024 09:29



Tgt Ion: 75 Resp: 16531
Ion Ratio Lower Upper
75 100
155 75.6 48.6 145.8
157 93.6 62.7 188.2





#93

1,2,4-Trichlorobenzene

Concen: 48.641 ug/l

RT: 14.925 min Scan# 2166

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:180 Resp: 127439

Ion Ratio Lower Upper

180 100

182 95.8 47.9 143.8





145 32.6 14.2 42.6

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024

Supervised By :Mahesh Dadoda 10/29/2024



#94

Hexachlorobutadiene

Concen: 48.490 ug/l

RT: 15.029 min Scan# 2183

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29




Tgt Ion:225 Resp: 70632


Ion Ratio Lower Upper

225 100

223 64.5 30.4 91.2

227 64.8 31.9 95.9





#95

Naphthalene

Concen: 46.315 ug/l

RT: 15.151 min Scan# 2201

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Instrument :

MSVOA_Y

ClientSampleId :




VSTDCCC050

Tgt Ion:128 Resp: 228860
 Ion Ratio Lower Upper
 128 100
 127 13.2 10.6 16.0
 129 10.5 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 10/29/2024
 Supervised By :Mahesh Dadoda 10/29/2024



#96

1,2,3-Trichlorobenzene

Concen: 47.596 ug/l

RT: 15.334 min Scan# 2233

Delta R.T. 0.006 min

Lab File: VY020029.D

Acq: 28 Oct 2024 09:29

Tgt Ion:180 Resp: 105414
 Ion Ratio Lower Upper
 180 100
 182 92.5 47.6 142.9
 145 34.2 15.4 46.1

