



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102824\
Data File : VY020028.D
Acq On : 28 Oct 2024 08:55
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Title : SW846 8260
Last Update : Wed Oct 16 05:44:48 2024



AutoFind: Scans 1753, 1754, 1755; Background Corrected with Scan 1745

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.5	47640	PASS
75	95	30	60	57.6	111773	PASS
95	95	100	100	100.0	194197	PASS
96	95	5	9	6.6	12841	PASS
173	174	0.00	2	1.2	1678	PASS
174	95	50	100	72.0	139784	PASS
175	174	5	9	7.9	11045	PASS
176	174	95	101	95.8	133875	PASS
177	176	5	9	6.6	8812	PASS