

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102825\
 Data File : VY023599.D
 Acq On : 28 Oct 2025 09:30
 Operator : SY/MD
 Sample : VY1028SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1028SBL01

Quant Time: Oct 29 06:21:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	751348	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1250620	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1019446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	391368	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	318592	50.708	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.420%
35) Dibromofluoromethane	7.634	113	383941	49.963	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.920%
50) Toluene-d8	10.109	98	1396250	48.790	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.580%
62) 4-Bromofluorobenzene	12.408	95	372675	39.444	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102825\
Data File : VY023599.D
Acq On : 28 Oct 2025 09:30
Operator : SY/MD
Sample : VY1028SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1028SBL01

Quant Time: Oct 29 06:21:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

