

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102925\
 Data File : VY023611.D
 Acq On : 29 Oct 2025 09:12
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 30 03:12:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.360	0.310	13.9	77	0.00
3 P	Chloromethane	0.490	0.444	9.4	79	0.00
4 C	Vinyl Chloride	0.639	0.605	5.3#	81	0.00
5 T	Bromomethane	0.536	0.500	6.7	85	0.00
6 T	Chloroethane	0.435	0.444	-2.1	85	0.00
7 T	Trichlorofluoromethane	0.887	0.903	-1.8	86	0.00
8 T	Diethyl Ether	0.232	0.223	3.9	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.495	-7.6	90	0.00
10 T	Methyl Iodide	0.512	0.569	-11.1	84	0.00
11 T	Tert butyl alcohol	0.031	0.028	9.7	90	0.00
12 CM	1,1-Dichloroethene	0.446	0.456	-2.2#	85	0.00
13 T	Acrolein	0.044	0.009	79.5#	16#	0.00
14 T	Allyl chloride	0.544	0.572	-5.1	87	0.00
15 T	Acrylonitrile	0.089	0.088	1.1	86	0.00
16 T	Acetone	0.100	0.099	1.0	95	0.00
17 T	Carbon Disulfide	1.277	1.219	4.5	79	0.00
18 T	Methyl Acetate	0.261	0.301	-15.3	98	0.00
19 T	Methyl tert-butyl Ether	1.072	1.083	-1.0	85	0.00
20 T	Methylene Chloride	0.538	0.510	5.2	89	0.00
21 T	trans-1,2-Dichloroethene	0.491	0.518	-5.5	88	0.00
22 T	Diisopropyl ether	1.202	1.327	-10.4	91	0.00
23 T	Vinyl Acetate	0.651	0.645	0.9	82	0.00
24 P	1,1-Dichloroethane	0.796	0.867	-8.9	91	0.00
25 T	2-Butanone	0.122	0.116	4.9	88	0.00
26 T	2,2-Dichloropropane	0.731	0.806	-10.3	91	0.00
27 T	cis-1,2-Dichloroethene	0.567	0.621	-9.5	91	0.00
28 T	Bromochloromethane	0.297	0.281	5.4	84	0.00
29 T	Tetrahydrofuran	0.069	0.065	5.8	82	0.00
30 C	Chloroform	0.875	0.975	-11.4#	93	0.00
31 T	Cyclohexane	0.696	0.685	1.6	85	0.00
32 T	1,1,1-Trichloroethane	0.795	0.868	-9.2	90	0.00
33 S	1,2-Dichloroethane-d4	0.418	0.381	8.9	78	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	79	0.00
35 S	Dibromofluoromethane	0.307	0.306	0.3	81	0.00
36 T	1,1-Dichloropropene	0.407	0.456	-12.0	90	0.00
37 T	Ethyl Acetate	0.161	0.152	5.6	79	0.00
38 T	Carbon Tetrachloride	0.493	0.553	-12.2	90	0.00
39 T	Methylcyclohexane	0.530	0.575	-8.5	86	0.00
40 TM	Benzene	1.288	1.443	-12.0	90	0.00
41 T	Methacrylonitrile	0.087	0.094	-8.0	93	0.00
42 TM	1,2-Dichloroethane	0.332	0.356	-7.2	89	0.00
43 T	Isopropyl Acetate	0.321	0.314	2.2	82	0.00
44 TM	Trichloroethene	0.377	0.412	-9.3	88	0.00
45 C	1,2-Dichloropropane	0.285	0.323	-13.3#	93	0.00
46 T	Dibromomethane	0.180	0.194	-7.8	88	0.00
47 T	Bromodichloromethane	0.453	0.510	-12.6	91	0.00
48 T	Methyl methacrylate	0.151	0.153	-1.3	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102925\
 Data File : VY023611.D
 Acq On : 29 Oct 2025 09:12
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 30 03:12:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	79	0.00
50 S	Toluene-d8	1.144	1.115	2.5	78	0.00
51 T	4-Methyl-2-Pentanone	0.166	0.170	-2.4	84	0.00
52 CM	Toluene	0.836	0.950	-13.6#	90	0.00
53 T	t-1,3-Dichloropropene	0.394	0.424	-7.6	87	0.00
54 T	cis-1,3-Dichloropropene	0.465	0.507	-9.0	88	0.00
55 T	1,1,2-Trichloroethane	0.242	0.267	-10.3	91	0.00
56 T	Ethyl methacrylate	0.279	0.294	-5.4	83	0.00
57 T	1,3-Dichloropropane	0.383	0.417	-8.9	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.129	-3.2	79	0.00
59 T	2-Hexanone	0.119	0.121	-1.7	85	0.00
60 T	Dibromochloromethane	0.339	0.373	-10.0	90	0.00
61 T	1,2-Dibromoethane	0.231	0.247	-6.9	87	0.00
62 S	4-Bromofluorobenzene	0.378	0.380	-0.5	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.516	0.565	-9.5	86	0.00
65 PM	Chlorobenzene	1.087	1.199	-10.3	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.432	-11.3	92	0.00
67 C	Ethyl Benzene	1.760	2.037	-15.7#	91	0.00
68 T	m/p-Xylenes	0.706	0.824	-16.7	91	0.00
69 T	o-Xylene	0.662	0.773	-16.8	91	0.00
70 T	Styrene	1.089	1.279	-17.4	91	0.00
71 P	Bromoform	0.240	0.252	-5.0	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.295	3.803	-15.4	92	0.00
74 T	N-amyl acetate	0.592	0.568	4.1	78	0.00
75 P	1,1,2,2-Tetrachloroethane	0.539	0.559	-3.7	90	0.00
76 T	1,2,3-Trichloropropane	0.441	0.413	6.3	76	0.00
77 T	Bromobenzene	0.876	0.941	-7.4	89	0.00
78 T	n-propylbenzene	3.813	4.478	-17.4	92	0.00
79 T	2-Chlorotoluene	2.233	2.539	-13.7	91	0.00
80 T	1,3,5-Trimethylbenzene	2.663	3.124	-17.3	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.180	0.179	0.6	85	0.00
82 T	4-Chlorotoluene	2.300	2.608	-13.4	92	0.00
83 T	tert-Butylbenzene	2.458	2.868	-16.7	91	0.00
84 T	1,2,4-Trimethylbenzene	2.657	3.115	-17.2	92	0.00
85 T	sec-Butylbenzene	3.527	4.161	-18.0	93	0.00
86 T	p-Isopropyltoluene	3.003	3.540	-17.9	92	0.00
87 T	1,3-Dichlorobenzene	1.706	1.860	-9.0	90	0.00
88 T	1,4-Dichlorobenzene	1.685	1.824	-8.2	90	0.00
89 T	n-Butylbenzene	2.616	3.084	-17.9	92	0.00
90 T	Hexachloroethane	0.646	0.723	-11.9	93	0.00
91 T	1,2-Dichlorobenzene	1.496	1.614	-7.9	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.088	0.082	6.8	79	0.00
93 T	1,2,4-Trichlorobenzene	0.932	0.955	-2.5	83	0.00
94 T	Hexachlorobutadiene	0.581	0.638	-9.8	88	0.00
95 T	Naphthalene	1.485	1.437	3.2	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102925\
Data File : VY023611.D
Acq On : 29 Oct 2025 09:12
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 30 03:12:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.808	0.813	-0.6	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6