

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103025\
 Data File : VY023624.D
 Acq On : 30 Oct 2025 11:53
 Operator : SY/MD
 Sample : Q3486-01
 Misc : 6.43g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP3

Quant Time: Oct 31 02:37:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	730570	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	1248092	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	997292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	374727	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	295297	48.337	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.680%
35) Dibromofluoromethane	7.640	113	379240	49.451	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.900%
50) Toluene-d8	10.109	98	1402224	49.098	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.200%
62) 4-Bromofluorobenzene	12.408	95	369288	39.165	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.340%
Target Compounds						
					Qvalue	
16) Acetone	3.879	43	8961	6.161	ug/l	94
20) Methylene Chloride	4.622	84	46945	5.969	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103025\
Data File : VY023624.D
Acq On : 30 Oct 2025 11:53
Operator : SY/MD
Sample : Q3486-01
Misc : 6.43g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP3

Quant Time: Oct 31 02:37:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

