

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103124\
 Data File : VY020094.D
 Acq On : 31 Oct 2024 15:02
 Operator : SY/MD
 Sample : P4597-07
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BLUE-2-1

Quant Time: Oct 31 17:13:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	28602	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	56175	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	52982	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	15736	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	27067	75.824	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	151.640%
35) Dibromofluoromethane	7.634	113	21205	59.372	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.740%
50) Toluene-d8	10.103	98	67870	47.740	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.480%
62) 4-Bromofluorobenzene	12.402	95	20630	44.676	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	89.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103124\
Data File : VY020094.D
Acq On : 31 Oct 2024 15:02
Operator : SY/MD
Sample : P4597-07
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BLUE-2-1

Quant Time: Oct 31 17:13:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

