

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103124\
 Data File : VY020102.D
 Acq On : 31 Oct 2024 18:09
 Operator : SY/MD
 Sample : P4616-03
 Misc : 6.22g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BP-F10-VOC

Quant Time: Nov 01 03:51:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	181210	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	381190	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	356911	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	136543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	133864	59.189	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.380%
35) Dibromofluoromethane	7.634	113	126495	52.193	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.380%
50) Toluene-d8	10.103	98	472649	48.994	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.980%
62) 4-Bromofluorobenzene	12.401	95	154068	49.168	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	98.340%
Target Compounds						
16) Acetone	3.878	43	3423	6.508	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103124\
Data File : VY020102.D
Acq On : 31 Oct 2024 18:09
Operator : SY/MD
Sample : P4616-03
Misc : 6.22g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BP-F10-VOC

Quant Time: Nov 01 03:51:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

