

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110124\
 Data File : VY020110.D
 Acq On : 01 Nov 2024 09:42
 Operator : SY/MD
 Sample : VY1101SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1101SBL01

Quant Time: Nov 01 16:02:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	193272	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	403931	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	365454	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	125829	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	138556	57.441	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.880%
35) Dibromofluoromethane	7.634	113	132476	51.584	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.160%
50) Toluene-d8	10.103	98	500171	48.928	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.860%
62) 4-Bromofluorobenzene	12.402	95	147112	44.305	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	88.620%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110124\
Data File : VY020110.D
Acq On : 01 Nov 2024 09:42
Operator : SY/MD
Sample : VY1101SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1101SBL01

Quant Time: Nov 01 16:02:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

