

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110124\
 Data File : VY020128.D
 Acq On : 01 Nov 2024 18:41
 Operator : SY/MD
 Sample : P4663-01
 Misc : 6.74g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TENNANT-PAD-2-3-STOCKPILE

Quant Time: Nov 04 00:51:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	159567	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	338625	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	324781	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	118518	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	121934	61.227	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.460%
35) Dibromofluoromethane	7.634	113	115825	53.798	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.600%
50) Toluene-d8	10.103	98	421310	49.162	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.320%
62) 4-Bromofluorobenzene	12.402	95	132684	47.667	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	95.340%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110124\
Data File : VY020128.D
Acq On : 01 Nov 2024 18:41
Operator : SY/MD
Sample : P4663-01
Misc : 6.74g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TENNANT-PAD-2-3-STOCKPILE

Quant Time: Nov 04 00:51:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

