

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110124\
 Data File : VY020129.D
 Acq On : 01 Nov 2024 19:04
 Operator : SY/MD
 Sample : P4663-03
 Misc : 6.25g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RES-BUILDING-PAD-NO-11

Quant Time: Nov 04 00:52:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	164371	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.609	114	344945	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	309515	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	105144	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	116799	56.935	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.860%
35) Dibromofluoromethane	7.634	113	113604	51.800	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.600%
50) Toluene-d8	10.103	98	426232	48.825	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.660%
62) 4-Bromofluorobenzene	12.401	95	124656	43.962	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	87.920%


Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110124\
Data File : VY020129.D
Acq On : 01 Nov 2024 19:04
Operator : SY/MD
Sample : P4663-03
Misc : 6.25g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RES-BUILDING-PAD-NO-11

Quant Time: Nov 04 00:52:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 98
Delta R.T. 0.006 min
Lab File: VY020129.D
Acq: 01 Nov 2024 19:04




Instrument :
MSVOA_Y
ClientSampleId :
RES-BUILDING-PAD-NO-11

Tgt Ion:168 Resp: 164371
Ion Ratio Lower Upper
168 100
99 64.3 39.1 58.7#



#33
1,2-Dichloroethane-d4
Concen: 56.935 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VY020129.D
Acq: 01 Nov 2024 19:04


Tgt Ion: 65 Resp: 116799
Ion Ratio Lower Upper
65 100
67 52.6 0.0 109.6








#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.609 min Scan# 1130
 Delta R.T. -0.006 min
 Lab File: VY020129.D
 Acq: 01 Nov 2024 19:04

Instrument : MSVOA_Y
 ClientSampleId : RES-BUILDING-PAD-NO-11







Tgt Ion:	114	Resp:	344945
Ion	Ratio	Lower	Upper
114	100		
63	22.5	0.0	35.0
88	14.8	0.0	27.2



#35
 Dibromofluoromethane
 Concen: 51.800 ug/l
 RT: 7.634 min Scan# 970
 Delta R.T. 0.000 min
 Lab File: VY020129.D
 Acq: 01 Nov 2024 19:04

Tgt Ion:	113	Resp:	113604
Ion	Ratio	Lower	Upper
113	100		
111	104.0	82.2	123.4
192	16.3	15.9	23.9





#50

Toluene-d8

Concen: 48.825 ug/l

RT: 10.103 min Scan# 1375

Delta R.T. -0.006 min

Lab File: VY020129.D

Acq: 01 Nov 2024 19:04

Instrument :

MSVOA_Y

ClientSampleId :





RES-BUILDING-PAD-NO-11

Tgt Ion: 98 Resp: 426232

Ion Ratio Lower Upper

98 100

100 64.9 52.0 78.0



#62

4-Bromofluorobenzene

Concen: 43.962 ug/l

RT: 12.401 min Scan# 1752

Delta R.T. -0.006 min

Lab File: VY020129.D

Acq: 01 Nov 2024 19:04




Tgt Ion: 95 Resp: 124656


Ion Ratio Lower Upper

95 100

174 73.5 0.0 175.6

176 71.8 0.0 171.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 1590

Delta R.T. -0.006 min

Lab File: VY020129.D

Acq: 01 Nov 2024 19:04

Instrument :

MSVOA_Y

ClientSampleId :

RES-BUILDING-PAD-NO-11





Tgt Ion:117 Resp: 309515

Ion Ratio Lower Upper

117 100

82 60.9 42.4 63.6

119 32.2 25.9 38.9



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VY020129.D

Acq: 01 Nov 2024 19:04

Tgt Ion:152 Resp: 105144

Ion Ratio Lower Upper

152 100

115 62.6 28.2 84.7

150 156.2 0.0 345.6

