

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110422\
 Data File : VY011314.D
 Acq On : 04 Nov 2022 21:40
 Operator : KP/MD
 Sample : N5436-05
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 250FERRYBLVD-TFS-05

Quant Time: Nov 07 01:56:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	245496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	405416	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	315773	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	112592	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	151872	57.843	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.680%
35) Dibromofluoromethane	7.716	113	153432	58.904	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	117.800%
50) Toluene-d8	10.179	98	419741	41.504	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	83.000%
62) 4-Bromofluorobenzene	12.477	95	144598	41.185	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.380%
Target Compounds						Qvalue
20) Methylene Chloride	4.710	84	46475	10.351	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110422\
Data File : VY011314.D
Acq On : 04 Nov 2022 21:40
Operator : KP/MD
Sample : N5436-05
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
250FERRYBLVD-TFS-05

Quant Time: Nov 07 01:56:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

