

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110422\
 Data File : VY011315.D
 Acq On : 04 Nov 2022 22:03
 Operator : KP/MD
 Sample : N5436-06
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 250FERRYBLVD-TFS-06

Quant Time: Nov 07 01:56:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	115605	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	191266	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	145762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	47775	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	79674	64.440	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 128.880%			
35) Dibromofluoromethane	7.716	113	77676	63.209	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 126.420%			
50) Toluene-d8	10.179	98	184123	38.590	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 77.180%			
62) 4-Bromofluorobenzene	12.477	95	62088	37.484	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 74.960%			
Target Compounds						
17) Carbon Disulfide	4.192	76	5106	1.236	ug/l #	94
20) Methylene Chloride	4.716	84	22645	10.904	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110422\
Data File : VY011315.D
Acq On : 04 Nov 2022 22:03
Operator : KP/MD
Sample : N5436-06
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
250FERRYBLVD-TFS-06

Quant Time: Nov 07 01:56:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

