

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110422\
 Data File : VY011317.D
 Acq On : 04 Nov 2022 22:48
 Operator : KP/MD
 Sample : N5436-08
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 250FERRYBLVD-TFS-08

Quant Time: Nov 07 01:56:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	235796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	394345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	302483	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	115426	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	153496	60.866	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 121.740%		
35) Dibromofluoromethane	7.716	113	155347	61.313	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 122.620%		
50) Toluene-d8	10.179	98	396802	40.337	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 80.680%		
62) 4-Bromofluorobenzene	12.477	95	146770	42.977	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 85.960%		
Target Compounds						Qvalue
16) Acetone	3.948	43	9340	12.269	ug/l #	83
20) Methylene Chloride	4.710	84	50956	12.607	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110422\
Data File : VY011317.D
Acq On : 04 Nov 2022 22:48
Operator : KP/MD
Sample : N5436-08
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
250FERRYBLVD-TFS-08

Quant Time: Nov 07 01:56:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

