

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110424\
 Data File : VY020148.D
 Acq On : 04 Nov 2024 17:20
 Operator : SY/MD
 Sample : P4694-03
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 Z-03A-VOC

Quant Time: Nov 05 00:35:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	28220	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.622	114	38712	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	12953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	1019	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	4417	12.541	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	25.080%#
35) Dibromofluoromethane	7.634	113	8015	32.564	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	65.120%
50) Toluene-d8	10.109	98	36805	37.567	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	75.140%
62) 4-Bromofluorobenzene	12.401	95	2295	7.212	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	14.420%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110424\
Data File : VY020148.D
Acq On : 04 Nov 2024 17:20
Operator : SY/MD
Sample : P4694-03
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
Z-03A-VOC

Quant Time: Nov 05 00:35:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

