

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110424\
 Data File : VY020151.D
 Acq On : 04 Nov 2024 18:30
 Operator : SY/MD
 Sample : P4693-07
 Misc : 5.40g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BP-G4-WC-VOC

Quant Time: Nov 05 00:36:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	154165	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	326420	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	310246	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	112462	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	123409	64.139	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.280%	
35) Dibromofluoromethane	7.640	113	110643	53.313	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.620%	
50) Toluene-d8	10.109	98	413424	50.046	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.100%	
62) 4-Bromofluorobenzene	12.408	95	130189	48.519	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 97.040%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110424\
Data File : VY020151.D
Acq On : 04 Nov 2024 18:30
Operator : SY/MD
Sample : P4693-07
Misc : 5.40g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BP-G4-WC-VOC

Quant Time: Nov 05 00:36:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

