

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
 Data File : VY006650.D
 Acq On : 05 Nov 2021 12:29
 Operator : SY/MD
 Sample : M4497-04
 Misc : 5.23G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-203

Quant Time: Nov 06 20:37:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	109113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	185524	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	170670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	66769	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	76549	57.011	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.020%
35) Dibromofluoromethane	7.728	113	57197	55.313	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.620%
50) Toluene-d8	10.185	98	238220	54.344	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.680%
62) 4-Bromofluorobenzene	12.483	95	71302	48.287	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.580%
Target Compounds					Qvalue	
16) Acetone	3.960	43	3540	9.156	ug/l	91
20) Methylene Chloride	4.735	84	8977	3.754	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
Data File : VY006650.D
Acq On : 05 Nov 2021 12:29
Operator : SY/MD
Sample : M4497-04
Misc : 5.23G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-203

Quant Time: Nov 06 20:37:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

