

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
 Data File : VY006653.D
 Acq On : 05 Nov 2021 13:40
 Operator : SY/MD
 Sample : M4498-04
 Misc : 5.03G/10ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-2-MD

Quant Time: Nov 06 20:38:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	119388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	201342	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	185204	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	75029	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	87188	59.346	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.700%
35) Dibromofluoromethane	7.728	113	62004	55.251	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.500%
50) Toluene-d8	10.185	98	256382	53.893	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.780%
62) 4-Bromofluorobenzene	12.483	95	78994	49.293	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
Data File : VY006653.D
Acq On : 05 Nov 2021 13:40
Operator : SY/MD
Sample : M4498-04
Misc : 5.03G/10ML/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-2-MD

Quant Time: Nov 06 20:38:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

