

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
 Data File : VY006654.D
 Acq On : 05 Nov 2021 14:03
 Operator : SY/MD
 Sample : M4498-07
 Misc : 5.05G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-2-LS

Quant Time: Nov 06 20:38:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	11043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	15857	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	12505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	3375	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	5923	43.586	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.180%
35) Dibromofluoromethane	7.722	113	2208	24.982	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	49.960%#
50) Toluene-d8	10.185	98	19812	52.879	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.760%
62) 4-Bromofluorobenzene	12.483	95	4062	32.185	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	64.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
Data File : VY006654.D
Acq On : 05 Nov 2021 14:03
Operator : SY/MD
Sample : M4498-07
Misc : 5.05G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-2-LS

Quant Time: Nov 06 20:38:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

