

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
 Data File : VY006655.D
 Acq On : 05 Nov 2021 14:45
 Operator : SY/MD
 Sample : M4501-01
 Misc : 3.72G/10ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11929

Quant Time: Nov 06 20:38:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	114891	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	196037	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	180293	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	72989	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	86489	61.174	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.340%
35) Dibromofluoromethane	7.728	113	59724	54.660	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.320%
50) Toluene-d8	10.185	98	249395	53.843	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.680%
62) 4-Bromofluorobenzene	12.483	95	75419	48.336	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
Data File : VY006655.D
Acq On : 05 Nov 2021 14:45
Operator : SY/MD
Sample : M4501-01
Misc : 3.72G/10ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11929

Quant Time: Nov 06 20:38:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

