

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
 Data File : VY006656.D
 Acq On : 05 Nov 2021 15:08
 Operator : SY/MD
 Sample : M4477-03
 Misc : 5.05G/5ML/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 T1-EP03-S

Quant Time: Nov 06 20:39:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	117396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	199162	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	190360	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	74941	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	85413	59.124	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.240%
35) Dibromofluoromethane	7.728	113	62074	55.919	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.840%
50) Toluene-d8	10.185	98	257071	54.629	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.260%
62) 4-Bromofluorobenzene	12.483	95	77033	48.596	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
Data File : VY006656.D
Acq On : 05 Nov 2021 15:08
Operator : SY/MD
Sample : M4477-03
Misc : 5.05G/5ML/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
T1-EP03-S

Quant Time: Nov 06 20:39:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

