

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
 Data File : VY006658.D
 Acq On : 05 Nov 2021 15:56
 Operator : SY/MD
 Sample : M4477-01
 Misc : 4.99G/5ML/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 T1-EP01-N

Quant Time: Nov 06 20:39:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	116563	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	199838	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	182436	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	72190	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	81233	56.633	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 113.260%	
35) Dibromofluoromethane	7.728	113	60234	54.078	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.160%	
50) Toluene-d8	10.185	98	255452	54.101	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.200%	
62) 4-Bromofluorobenzene	12.483	95	76060	47.820	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 95.640%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
Data File : VY006658.D
Acq On : 05 Nov 2021 15:56
Operator : SY/MD
Sample : M4477-01
Misc : 4.99G/5ML/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
T1-EP01-N

Quant Time: Nov 06 20:39:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

