

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
 Data File : VY006665.D
 Acq On : 05 Nov 2021 18:41
 Operator : SY/MD
 Sample : M4477-10
 Misc : 5.06G/10ML/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 T5-B1-NE

Quant Time: Nov 06 20:42:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	110093	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	190111	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	179735	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	79022	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	78628	58.038	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.080%
35) Dibromofluoromethane	7.728	113	58044	54.778	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.560%
50) Toluene-d8	10.185	98	248833	55.396	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.800%
62) 4-Bromofluorobenzene	12.483	95	82545	54.552	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110521\
Data File : VY006665.D
Acq On : 05 Nov 2021 18:41
Operator : SY/MD
Sample : M4477-10
Misc : 5.06G/10ML/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
T5-B1-NE

Quant Time: Nov 06 20:42:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

