



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110625\
Data File : VY023691.D
Acq On : 06 Nov 2025 08:09
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
Title : SW846 8260
Last Update : Wed Nov 05 04:41:50 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	31323	PASS
75	95	30	60	50.0	90808	PASS
95	95	100	100	100.0	181653	PASS
96	95	5	9	6.5	11780	PASS
173	174	0.00	2	1.3	2185	PASS
174	95	50	100	91.0	165333	PASS
175	174	5	9	7.6	12503	PASS
176	174	95	101	97.8	161725	PASS
177	176	5	9	6.6	10598	PASS