

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110722\
 Data File : VY011329.D
 Acq On : 07 Nov 2022 12:33
 Operator : KP/MD
 Sample : N5459-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-03-110322

Quant Time: Nov 08 03:28:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	253999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	429723	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	375096	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	153543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	148980	54.842	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.680%	
35) Dibromofluoromethane	7.710	113	154292	55.884	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.760%	
50) Toluene-d8	10.179	98	508026	47.392	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 94.780%	
62) 4-Bromofluorobenzene	12.477	95	173776	46.696	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 93.400%	
Target Compounds						Qvalue
20) Methylene Chloride	4.710	84	74846	19.223	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110722\
Data File : VY011329.D
Acq On : 07 Nov 2022 12:33
Operator : KP/MD
Sample : N5459-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-03-110322

Quant Time: Nov 08 03:28:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

