

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110722\
 Data File : VY011334.D
 Acq On : 07 Nov 2022 14:50
 Operator : KP/MD
 Sample : N5495-04
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-31A

Quant Time: Nov 08 03:29:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	315648	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	512628	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	421626	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	155398	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	180293	53.406	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.820%	
35) Dibromofluoromethane	7.710	113	190898	57.960	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 115.920%	
50) Toluene-d8	10.179	98	640839	50.113	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.220%	
62) 4-Bromofluorobenzene	12.483	95	198041	44.610	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 89.220%	
Target Compounds						
20) Methylene Chloride	4.704	84	47226	7.008	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110722\
Data File : VY011334.D
Acq On : 07 Nov 2022 14:50
Operator : KP/MD
Sample : N5495-04
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-31A

Quant Time: Nov 08 03:29:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

