

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110722\
 Data File : VY011345.D
 Acq On : 07 Nov 2022 19:04
 Operator : KP/MD
 Sample : N5452-01
 Misc : 7.48g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL-PILE

Quant Time: Nov 08 03:31:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	268783	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	448056	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	370362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	139637	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	147643	51.360	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.720%
35) Dibromofluoromethane	7.710	113	160483	55.748	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.500%
50) Toluene-d8	10.179	98	501426	44.862	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.720%
62) 4-Bromofluorobenzene	12.477	95	155747	40.139	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110722\
Data File : VY011345.D
Acq On : 07 Nov 2022 19:04
Operator : KP/MD
Sample : N5452-01
Misc : 7.48g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOIL-PILE

Quant Time: Nov 08 03:31:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

