

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110725\
 Data File : VY023709.D
 Acq On : 07 Nov 2025 15:29
 Operator : SY/MD
 Sample : Q3580-01
 Misc : 5.71g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 346

Quant Time: Nov 08 03:28:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 05 04:41:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	762740	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1105415	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	690272	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	205694	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	181018	27.822	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	55.640%
35) Dibromofluoromethane	7.634	113	274375	39.718	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.440%
50) Toluene-d8	10.109	98	1155674	44.910	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.820%
62) 4-Bromofluorobenzene	12.408	95	227125	27.292	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	54.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110725\
Data File : VY023709.D
Acq On : 07 Nov 2025 15:29
Operator : SY/MD
Sample : Q3580-01
Misc : 5.71g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
346

Quant Time: Nov 08 03:28:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 05 04:41:50 2025
Response via : Initial Calibration

