



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110725\
Data File : VY023703.D
Acq On : 07 Nov 2025 08:19
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
Title : SW846 8260
Last Update : Wed Nov 05 04:41:50 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.5	30512	PASS
75	95	30	60	49.8	86981	PASS
95	95	100	100	100.0	174805	PASS
96	95	5	9	6.4	11143	PASS
173	174	0.00	2	1.3	2081	PASS
174	95	50	100	90.5	158173	PASS
175	174	5	9	7.8	12329	PASS
176	174	95	101	96.0	151837	PASS
177	176	5	9	6.8	10310	PASS