

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110822\
 Data File : VY011372.D
 Acq On : 08 Nov 2022 19:05
 Operator : KP/MD
 Sample : N5436-02RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 250FERRYBLVD-TFS-02RE

Quant Time: Nov 09 03:35:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	7296	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	11801	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	9225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	2698	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	5316	68.127	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.260%
35) Dibromofluoromethane	7.715	113	3342	44.078	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.160%
50) Toluene-d8	10.178	98	13077	44.422	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.840%
62) 4-Bromofluorobenzene	12.477	95	3749	36.684	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	73.360%
Target Compounds						
20) Methylene Chloride	4.710	84	1960	17.031	ug/l	# 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110822\
Data File : VY011372.D
Acq On : 08 Nov 2022 19:05
Operator : KP/MD
Sample : N5436-02RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
250FERRYBLVD-TFS-02RE

Quant Time: Nov 09 03:35:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

