

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110822\
 Data File : VY011373.D
 Acq On : 08 Nov 2022 19:27
 Operator : KP/MD
 Sample : N5436-04RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 250FERRYBLVD-TFS-04RE

Quant Time: Nov 09 03:35:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	29984	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	45585	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	39758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	16359	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	19693	61.410	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.820%	
35) Dibromofluoromethane	7.716	113	20056	68.478	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 136.960%	
50) Toluene-d8	10.179	98	54087	47.564	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 95.120%	
62) 4-Bromofluorobenzene	12.477	95	16941	42.914	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 85.820%	
Target Compounds						Qvalue
20) Methylene Chloride	4.710	84	5688	10.384	ug/l	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110822\
Data File : VY011373.D
Acq On : 08 Nov 2022 19:27
Operator : KP/MD
Sample : N5436-04RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
250FERRYBLVD-TFS-04RE

Quant Time: Nov 09 03:35:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

