


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110822\
Data File : VY011350.D
Acq On : 08 Nov 2022 09:42
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Title : SW846 8260
Last Update : Tue Nov 01 11:56:18 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.5	22248	PASS
75	95	30	60	47.9	68821	PASS
95	95	100	100	100.0	143717	PASS
96	95	5	9	6.8	9721	PASS
173	174	0.00	2	1.0	1197	PASS
174	95	50	100	86.2	123875	PASS
175	174	5	9	7.3	9029	PASS
176	174	95	101	96.3	119264	PASS
177	176	5	9	6.7	8023	PASS