

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111025\
 Data File : VY023721.D
 Acq On : 10 Nov 2025 17:44
 Operator : SY/MD
 Sample : Q3597-01
 Misc : 6.27g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1

Quant Time: Nov 11 02:10:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	414776	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	687106	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	635249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	305577	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	223010	64.297	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.600%
35) Dibromofluoromethane	7.640	113	245335	58.110	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.220%
50) Toluene-d8	10.109	98	923004	58.704	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	117.400%
62) 4-Bromofluorobenzene	12.407	95	311039	59.920	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	119.840%
Target Compounds						
						Qvalue
16) Acetone	3.878	43	7960	9.640	ug/l	93
20) Methylene Chloride	4.628	84	33861	7.584	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111025\
Data File : VY023721.D
Acq On : 10 Nov 2025 17:44
Operator : SY/MD
Sample : Q3597-01
Misc : 6.27g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1

Quant Time: Nov 11 02:10:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

