

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111025\
 Data File : VY023725.D
 Acq On : 10 Nov 2025 19:15
 Operator : SY/MD
 Sample : Q3587-01
 Misc : 5.43g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-08-110725

Quant Time: Nov 11 02:11:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	384683	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	605252	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	483827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	195201	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	188030	58.453	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.900%
35) Dibromofluoromethane	7.640	113	207184	55.710	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.420%
50) Toluene-d8	10.109	98	694270	50.128	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.260%
62) 4-Bromofluorobenzene	12.408	95	191297	41.836	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.680%
Target Compounds						Qvalue
20) Methylene Chloride	4.622	84	19285	4.657	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111025\
Data File : VY023725.D
Acq On : 10 Nov 2025 19:15
Operator : SY/MD
Sample : Q3587-01
Misc : 5.43g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-08-110725

Quant Time: Nov 11 02:11:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

