

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111122\
 Data File : VY011406.D
 Acq On : 11 Nov 2022 10:30
 Operator : KP/MD
 Sample : VY1111SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1111SBL01

Quant Time: Nov 11 13:32:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	245473	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	410416	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	367162	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	169940	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	107840	52.426	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.860%
35) Dibromofluoromethane	7.716	113	116425	50.618	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.240%
50) Toluene-d8	10.179	98	394592	44.907	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.820%
62) 4-Bromofluorobenzene	12.477	95	138292	45.123	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111122\
Data File : VY011406.D
Acq On : 11 Nov 2022 10:30
Operator : KP/MD
Sample : VY1111SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1111SBL01

Quant Time: Nov 11 13:32:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 10 17:47:26 2022
Response via : Initial Calibration

